

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.3
Project Manager: Daniel H. Ginsborg

Work Order: 22L1744
Received: 12/14/22 17:06
Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253122 Analyst: AP1										
Blank (2253122-BLK1) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK2) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK3) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK4) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK5) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK6) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK7) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK8) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253122-BLK9) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	ND	1.0	ug/L							
LCS (2253122-BS1) Prepared & Analyzed: 12/30/22										
Lead (Low Level)	49.2	1.0	ug/L	50		98	85-115			

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.3
Project Manager: Daniel H. Ginsborg

Work Order: 22L1744
Received: 12/14/22 17:06
Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253122 Analyst: AP1										
LCS (2253122-BS2)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	51.5	1.0	ug/L	50		103	85-115			
LCS (2253122-BS3)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	52.8	1.0	ug/L	50		106	85-115			
LCS (2253122-BS4)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	53.1	1.0	ug/L	50		106	85-115			
LCS (2253122-BS5)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	54.6	1.0	ug/L	50		109	85-115			
LCS (2253122-BS6)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	53.8	1.0	ug/L	50		108	85-115			
LCS (2253122-BS7)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	54.3	1.0	ug/L	50		109	85-115			
LCS (2253122-BS8)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	55.6	1.0	ug/L	50		111	85-115			
LCS (2253122-BS9)				Prepared & Analyzed: 12/30/22						
Lead (Low Level)	55.5	1.0	ug/L	50		111	85-115			
Matrix Spike (2253122-MS1)				Source: 22L1539-01		Prepared & Analyzed: 12/30/22				
Lead (Low Level)	50.1	1.0	ug/L	50	0.627	99	70-130			
Matrix Spike (2253122-MS2)				Source: 22L1539-11		Prepared & Analyzed: 12/30/22				
Lead (Low Level)	52.5	1.0	ug/L	50	0.230	104	70-130			

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.3
Project Manager: Daniel H. Ginsborg

Work Order: 22L1744
Received: 12/14/22 17:06
Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253122 Analyst: AP1										
Matrix Spike (2253122-MS3) Source: 22L1542-10 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	52.9	1.0	ug/L	50	1.62	103	70-130			
Matrix Spike (2253122-MS4) Source: 22L1742-05 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	52.5	1.0	ug/L	50	ND	105	70-130			
Matrix Spike (2253122-MS5) Source: 22L1744-05 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	56.2	1.0	ug/L	50	2.29	108	70-130			
Matrix Spike (2253122-MS6) Source: 22L1801-06 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	56.5	1.0	ug/L	50	0.572	112	70-130			
Matrix Spike (2253122-MS7) Source: 22L1801-16 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	56.0	1.0	ug/L	50	0.530	111	70-130			
Matrix Spike Dup (2253122-MSD1) Source: 22L1539-01 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	50.2	1.0	ug/L	50	0.627	99	70-130	0.3	30	
Matrix Spike Dup (2253122-MSD2) Source: 22L1539-11 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	51.6	1.0	ug/L	50	0.230	103	70-130	2	30	
Matrix Spike Dup (2253122-MSD3) Source: 22L1542-10 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	53.8	1.0	ug/L	50	1.62	104	70-130	2	30	
Matrix Spike Dup (2253122-MSD4) Source: 22L1742-05 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	52.5	1.0	ug/L	50	ND	105	70-130	0.02	30	
Matrix Spike Dup (2253122-MSD5) Source: 22L1744-05 Prepared & Analyzed: 12/30/22										
Lead (Low Level)	55.7	1.0	ug/L	50	2.29	107	70-130	0.7	30	

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.3
Project Manager: Daniel H. Ginsborg

Work Order: 22L1744
Received: 12/14/22 17:06
Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-----------

Batch 2253122

Analyst: AP1

Matrix Spike Dup (2253122-MSD6)

Source: 22L1801-06

Prepared & Analyzed: 12/30/22

Lead (Low Level)	55.4	1.0 ug/L		50	0.572	110	70-130	2	30	
------------------	------	----------	--	----	-------	-----	--------	---	----	--

Matrix Spike Dup (2253122-MSD7)

Source: 22L1801-16

Prepared & Analyzed: 12/30/22

Lead (Low Level)	55.6	1.0 ug/L		50	0.530	110	70-130	0.8	30	
------------------	------	----------	--	----	-------	-----	--------	-----	----	--

J Detected below the Reporting Limit; reported concentration is estimated; (J-Flag)

ND Analyte NOT DETECTED at or above the reporting limit

Jeanette Hernandez

Project Manager